

Guidelines for the Blood Transfusion Services

Chapter 11: Reagent manufacture

<http://www.transfusionguidelines.org/red-book/chapter-11-reagent-manufacture>

Chapter 11:

Reagent manufacture

11.1: Guidelines for reagent manufacture

11.1.1: Introduction

All reagents used to determine the group of human red cells and to detect red cell antibodies must comply with Medical Devices Regulations 2002 (SI 2002 No 618, as amended) (UK MDR 2002) and all associated standards.

General guidelines for reagent manufacture are presented in this section. In other sections additional guidelines are given for particular reagents.

Where specific reference to British Standard European Standard (BS EN) documents is given this is the most recent version. It is intended that these guidelines refer to the current requirements contained in the applicable documents so the phrase 'and subsequent revisions' should be assumed whenever a specific reference is given.

11.1.2: Reference preparations

Reference preparations for use with these guidelines can be found on the National Institute for Biological Standards and Control (NIBSC) website at www.nibsc.org.

See section 11.3 for further information.

11.1.3: Definitions

Antibody identification is a test or combination of tests designed to determine the specificity of atypical antibodies.

Antibody screening is a test or combination of tests designed to detect atypical antibodies.

A **batch** of reagent is a defined quantity of material or of bulk, intermediate or finished product that is intended or purported to be uniform in character and quality, and which has been produced during a defined cycle of manufacture. A batch may be divided into sub-batches. A batch is sometimes described as a 'lot'.

A **batch of tests** is defined as a number of tests set up at the same time, under the same conditions and processed in a similar manner.

A **blood grouping kit** comprises a set of blood grouping components (reagents or materials), packaged together, intended by the manufacturer to be used together for determining one or more blood groups.

A **blood grouping reagent** is a reagent, used alone or in combination with other materials, intended by the manufacturer for the determination of a blood group of an individual.

- A blood grouping reagent recommended by the manufacturer for the detection of A (i.e. subgroups A₁ and A₂) A_x and B should be named **anti-A,B blood grouping reagent**.

A **blood grouping system** is an *in vitro* diagnostic medical device intended by the manufacturer to be used for determining one or more blood groups.

Clinically important or clinically significant antibody is a red cell antibody which will produce significantly accelerated red cell destruction when combined *in vivo* with its corresponding antigen.

Expiry date is the date beyond which performance of the reagent cannot be assured and is based upon the stability of the reagent.

Fresh serum for complement activity stored in the liquid state should be used within 8 hours of donation. When used after storage at –70°C or below, the 8-hour liquid storage period refers to the time both before and after frozen storage. Unless validated, the maximum period of frozen storage shall be 6 months at this temperature.

An **immediate container** is a medium adequate to protect the content(s) from contamination and/or physical damage. For example, a sealed vial, ampoule or bottle, a foiled pouch or a sealed plastic bag.

Atypical blood group antibodies are those of non-ABO specificity.

The **manufacturer** is a natural or legal person who manufactures or fully refurbishes a device or has a device designed, manufactured or fully refurbished, and markets that device under its name or trademark.

The name for a blood grouping reagent derived from monoclonal materials should include the word **monoclonal**.

A **monospecific blood grouping reagent** is one containing an antibody or blend of antibodies specific for one antigen, e.g. anti-A, anti-IgG.

A **polyspecific blood grouping reagent** is one containing a blend of antibodies specific for more than one antigen.

Polyspecific anti-human globulin reagent should be the name for a reagent which contains anti-human IgG and anti-human complement (C3d) activity, and is recommended by the manufacturer for use in both the direct and indirect anti-human globulin techniques, i.e. for the detection of red cell bound human IgG, and C3 complement in the form EiC3b and EC3d irrespective of the presence of other human immunoglobulin or human complement specificities.

Potency titre is a term used to describe the highest dilution of a reagent that effects a grade 2 endpoint reaction in tube or a grade 1 endpoint in column agglutination technologies.

Prozone is the term used to denote the absence or weakening of agglutination with excess of antibody.

A **reagent control** is a reagent made to the same formulation as a blood grouping reagent but without the specific blood group antibody reactivity. If the reagent control contains serum or plasma, the reagent control should be shown to be free from specific blood group antibody reactivity.

A **reference preparation** is prepared nationally or locally and contains a known or agreed concentration of the activity being measured. It should be assayed to establish the sensitivity or calibration of a test procedure or reagent.

Sensitivity in relation to these guidelines is a term defining the limit of detectable specific reactions using reagents or test systems. These guidelines specify levels of sensitivity that should be achieved.

Shelf life is the period until expiry date.

Specificity in relation to these guidelines is a term defining the ability of a reagent or test system to react selectively. In particular terms, it represents the absence of unwanted or false-positive reactions.

Test monitors are a series of samples included as part of each batch of tests, which provide part of the release algorithm for a batch of tests.

Validation is the confirmation, through the provision of objective evidence, that the requirements for a specific or intended use have been fulfilled. Validation of a manufacturing method is to ensure that the product will be of the quality required for its intended use and that tests used in monitoring will accurately reflect the quality of the product.

Verification is the confirmation, through the provision of objective evidence, that specific requirements have been fulfilled.

Undiluted in these guidelines means the reagent as intended for use by the manufacturer. This term includes a diluted reagent if the reagent is supplied in a form requiring dilution by the user prior to use, as specified in the manufacturer's 'instructions for use'.

An **unequivocal** reaction in a test system is a reaction that is unambiguous. In the manual tube test, this is defined as a reaction of grade 3 or greater. In column tests this is defined as a 1+ or greater reaction.

11.1.4: General manufacturing considerations

11.1.4.1: Good manufacturing practice

Reagents for blood group serology must be manufactured in accordance with the Medical Devices Regulation 2002 (SI 2002 No 618) (UK MDR 2002) and all associated documents.

Guidance on the principles of good manufacturing practice can be obtained from *Rules and Guidance for Pharmaceutical Manufacturers and Distributors*.²

- The method of manufacture should result in a product within an immediate container that is homogeneous and free of properties which adversely affect its intended use throughout its recommended shelf life. The reagent should have no precipitate, particles or fibrin gel.
- Each batch or sub-batch should be specifically identified by a distinctive combination of numbers and /or letters (batch reference) which permits its history to be traced.
- Reagents should be produced by a validated process that is shown to be suitable for the intended purpose, including any methods for preserving red cells prior to their preparation as reagent red cells.
- The manufacturer should monitor the batch-to-batch performance of the blood grouping reagent (e.g. by the reaction against some internal reference material) in order to provide consistency of

performance. This is particularly important when the blood grouping reagent is provided as a test system, kit or kit component, when the performance may be dependent on the characteristics of other system variables or kit components.

11.1.4.2: Risk management

Risk management should be performed in accordance with:

- BS EN ISO 14971:2019 Medical Devices – Application of Risk Management to Medical Devices.
- BS EN 13641:2002 Elimination or Reduction of Risk of Infection Related to *in vitro* Diagnostic Reagents.

11.1.4.3: Performance evaluation

Performance evaluation should be undertaken in accordance with:

- BS EN ISO 23640 *In vitro* diagnostic medical devices, Evaluation of stability of *in vitro* diagnostic reagents.
- Reagents must also comply with the Common Technical Specifications for *In Vitro* Diagnostic Medical Devices (where they exist).

11.1.4.4: Stability data

Stability testing should be performed in accordance with:

- BS EN ISO 23640 *In vitro* diagnostic medical devices, Evaluation of stability of *in vitro* diagnostic reagents.

11.1.4.5: Date of manufacture

- For blood grouping reagents the date of manufacture is the date of commencement of the last potency test on the batch or sub-batch that indicates attainment of the required specification.
- For reagent red cells the date of manufacture is the date of collection from the donor. Where reagent red cells are prepared from more than one donor, the date of collection of the oldest donation should be recorded as the date of manufacture.
- Where a freezing process is used to preserve red cells before their preparation for issue as reagent red cells, the date of manufacture is the date of recovery from the frozen state.

11.1.4.6: Colour coding of reagents

No colouring agent should be added to reagents for blood group serology except that:

- Polyspecific anti-human globulin reagents may be coloured green, anti-A may be coloured blue, anti-B may be coloured yellow.
- The colorant should not interfere with the observation of the test result.
- 'Bespoke' antisera for use on automation may be coloured providing the information contained in the barcode on each bottle contains sufficient identifiers (specificity and lot number) to provide assurance that the intended test has been performed. The colours used for other specificities should not be coloured blue or yellow to avoid confusion with those for anti-A and anti-B reagents.

11.1.4.7: Freedom from microbial contaminants

- Reagents should be prepared using validated processes to produce a final product free from microbial contaminants that adversely affect the unopened product during storage at the recommended temperature. The manufacturer should routinely monitor the efficacy of the process used in the manufacture of the reagent.
- A preservative may be included in the reagent to minimise the effects of contamination during use if the preservative has been shown not to adversely affect the product during storage or use.
- Other than reagent red cells, all reagents for blood group serology recommended by the manufacturer for storage in the liquid state, should be filtered through a sterile filter of pore size not exceeding 0.22 µm. All reagents should be dispensed into the immediate container under aseptic conditions.
- Tests for contamination do not give absolute assurance of freedom from microbial contaminants. Bactericidal agents in common use for blood grouping reagents do not guarantee the absence of microbial agents after opening of the container.

11.1.4.8: Retained samples

- A minimum of 1% or three immediate containers, whichever is less, of each batch of reagents other than reagent red cells should be retained and stored as recommended by the manufacturer to enable analysis of reported defects. Such samples should be retained for at least 6 months beyond the expiry date.
- A minimum of two final containers of each batch of reagent red cells should be retained and stored as recommended by the manufacturer to enable analysis of reported defects. Such samples should be retained for at least 10 days beyond the expiry date.

11.1.4.9: Tests required

The manufacturer should test, as described in these guidelines, each lot of a reagent obtained from the immediate container to be supplied for use (see section 11.2.1).

11.1.4.10: Human source material

Existing procedures in the UK Blood Transfusion Services for consent to donate are sufficient to allow cellular and plasma materials collected as part of the donation process to be used as reagents without further explicit consent.

Samples/donations that are obtained specifically for reagent purposes will require additional consenting of the donor, and must have appropriate ethical approval. Donor materials that are obtained and retained for genomic or nucleic acid testing must comply with the regulations laid down by The Human Tissue Act 2004 (except Scotland).³

Residual samples retained from patient testing laboratories may be used without further explicit consent, if anonymised.⁴ Additional samples taken from patients specifically for reagent use will require ethical approval and explicit consent. All patient samples acquired and retained must comply with the regulations laid down by the Human Tissue Act (2004).

Each individual donation or sample of human material in a reagent for blood group serology shall be tested and found negative for mandatory microbiological tests required by the UK Blood Transfusion Services for blood donations (see Chapter 9). A statement is required in the 'instructions for use' to this effect.

In case there is a need to perform retrospective microbiological testing on material used to prepare *in vitro* diagnostic devices, an archive sample collected at the same time and from the same donor(s) used to prepare the device should be taken and stored for up to six months.

11.1.4.11: Label requirements

The label must conform to the requirements of:

- BS EN 18113:2011 Information Supplied by the Manufacturer with *in vitro* Diagnostic Reagents for Professional Use.
- The label fixed to the immediate container of a reagent should leave uncovered sufficient area of the full length or circumference of the container to allow ready visual inspection of the contents.
- The specificity of the reagent for blood group serology should be of a print size which is clearly legible. The print size of other information on the label should not exceed that used for the specificity of the reagent.
- The typeface used should clearly differentiate between antigens and related antibody specificities represented by upper and lower-case characters, e.g. C/c, S/s and K/k.
- For products needing to be prepared in the final form by the user following the instructions of the manufacturer and to be retained in the manufacturer's immediate container, a space should be available on the container label for the user to write the expiry date of the prepared product when stored as recommended by the manufacturer.
- The main panel of labels of enzyme-treated reagent red cells may be coloured pink in order to be distinguishable from non-enzyme-treated reagent red cells. Pantone colour reference 223 is recommended.

For other reagents, any colour appearing on the main panel of the label should comply with Food and Drug Administration regulations (21 CFR 660.28) as shown in Table 11.1.

Table 11.1 Label colour coding

Specificity	Colour	Code
anti-A	Blue	305C
anti-B	Yellow	102C
anti-C	Pink	204C
anti-D	Grey	429C
anti-E	Brown	465C
anti-CDE	Orange	151C

anti-c	Lavender	529C
anti-e	Green	577C

11.1.4.12: Instructions for use (package insert)

The instructions for use must conform to the requirements of BS EN 18113:2011 Information Supplied by the Manufacturer with *in vitro* Diagnostic Reagents for Professional Use.

In addition:

- For blood grouping reagents containing monoclonal antibodies, the identity of the cell line(s) from which the monoclonal antibodies have been derived.
- For reagent red cells for antibody screening and for antibody identification, the 'antigen profile' of the component cell samples should have the lot number and expiry date of the reagent to which it refers.
- A statement that loss of reactivity may occur during the stated shelf life of the red cells and that since this loss is partly determined by characteristics of individual blood donations or donors, which cannot be predicted or controlled, the conditions of storage and use recommended by the manufacturer should be rigidly applied.
- For enzyme-treated reagent red cells, information should be given concerning those antigens which are rendered inactive or less active by the enzyme treatment used.

11.2: Specifications, performance evaluation and quality control of blood grouping reagents

11.2.1: Blood typing antisera

11.2.1.1: General requirements

- It is essential that blood grouping reagents are prepared using reliable manufacturing procedures that are consistently capable of producing safe and efficacious products. The products must comply with requirements of the EU Directive (98/79/EC) on *in vitro* diagnostic medical devices and other relevant international standards detailed in section 11.3.
- The term weak D is used in these guidelines to indicate a weakened expression of a normal D antigen. The term partial D is used in these recommendations to indicate the expression of only a part of the normal D antigen. The reactivity of RhD blood grouping reagents against partial D red cells is determined by the nature of the D variant, the anti-D reagent and the technique used.

- Red cell samples with partial antigen expression (e.g. partial D) or weak antigen expression (e.g. A_x) may not react with some reagents and, where this is known to be true, must be stated in the limitations.
- The blood grouping reagent is satisfactory if an unequivocal positive result is obtained with all the red cell samples having the antigen corresponding to the blood grouping reagent being assessed, by all the methods recommended for use by the manufacturer.
- If reactivity is claimed by the manufacturer against weak variants or subgroups of a particular antigen, red cells from at least two confirmed/reference samples should be tested (see Table 11.3).

The grading system shown in Table 11.2 is used throughout these guidelines for manual tube/microplate serological testing.

Table 11.2 Grading system for serological tests

Reaction grade	Description
Grade 5	Cell button remains in one clump or dislodges into a few large clumps
Grade 4	Cell button dislodges into numerous large clumps
Grade 3	Cell button dislodges into many small clumps
Grade 2	Cell button dislodges into finely granular but definite, small clumps
Grade 1	Cell button dislodges into fine granules
Grade 0	Negative result
Unless otherwise stated, an unequivocal manual tube reaction is defined as a grade 3 or greater and for column tests as 1+ or greater	

11.2.1.2: Performance evaluation

Performance evaluation should be undertaken in accordance with:

- BS EN 13612:2002 Performance Evaluation of *In Vitro* Diagnostic Medical Devices.
- Reagents listed in Annex II, List A, of the EU *In Vitro* Diagnostic Medical Devices Directive must also comply with the Common Technical Specifications for *In Vitro* Diagnostic Medical Devices (2009/108/EC).

Stability testing should be performed in accordance with BS EN ISO 23640 *in vitro* diagnostic medical devices. Evaluation of stability of *in vitro* diagnostic reagents.

Where appropriate, the following requirements should also be included in performance evaluation:

- In the case of polyclonal antibodies, contaminating antibodies to antigens having a prevalence of greater than 99% in the general population of the UK should be excluded. Negative results in tests using samples of red cells from four different individuals who lack the antigen corresponding to the antibody specificity under test. Tests for the presence of contaminating ABO antibodies should be performed with red cells from a minimum of two individuals of group A1 and two of group B who lack the antigen corresponding to the antibody specificity of the reagent but have the antigens to the potential contaminating antibodies should be obtained.
- If tests using all methods recommended for use by the manufacturer do not exclude the presence of antibodies to the following antigens, these antibody specificities should be stated in the package insert² as not having been excluded in specificity testing:
 - Xg^a, Do^a, Yt^b, Co^b, Wr^a and V^w.
- Blood grouping reagents which are chemically modified, and/or contain in their formulation a potentiator of agglutination, or require the user to add a potentiator, shall be tested, by all methods recommended by the manufacturer, with red cells lacking the antigen corresponding to the antibody specificity under test but sensitised with an IgG antibody to effect a grade 5 reaction in the anti-human globulin technique.
- Potentiated blood grouping reagents producing agglutination by those methods recommended by the manufacturer, should be supplied with a reagent control that has been shown to effect a degree of non-specific reaction with IgG-coated red cells similar to the corresponding blood grouping reagent.
- Blood grouping reagents recommended for use by a direct agglutination method should not contain antibodies reactive against red cells coated with IgG when used by direct agglutination methods recommended by the manufacturer.

11.2.1.3: Batch release testing requirements

Specificity tests

- The manufacturer must provide a certificate of analysis to customers once evidence has been obtained by the manufacturer that the product achieves the specificity and reactivity claimed by the manufacturer for each method recommended by the manufacturer, Assurance of Specificity should be determined in accordance with the requirements in Table 11.3. The certificate of analysis should also ensure that the potency of the material meets the requirements of the final bullet point on Potency below.
- If a range of incubation times or incubation temperatures is recommended by the manufacturer, the range(s) should be used in these test procedures.

Requirements

- Blood grouping reagents should not produce a positive reaction when tested with red cells lacking the antigen corresponding to the antibody specificity under test, by any method recommended for use by the manufacturer. Should reactivity to a low-frequency antigen be observed with subsequent batches of a reagent, this fact should be brought to the attention of all primary consignees of that

reagent.

- Rouleaux formation, prozone or haemolysis should not occur in tests using any of the methods recommended by the manufacturer.

Potency tests – tube or microplate methods

- Potency titrations should be performed in accordance with the manufacturer's recommended method of use using an appropriate diluent.
- Manufacturers should compare the potency titre of each batch of reagent with an appropriate reference preparation (see section 11.3).
- Potency titrations for each batch tested should equal or exceed any existing British or International reference preparations.

Table 11.3 Requirements for conventional blood typing reagents

Antibody specificity	Specification	Performance evaluation (as a minimum, two examples of the following reference cells should be included if available)*	Batch release testing					
			Specificity				Potency	
			Positive reactors		Negative reactors			
			Cell type	No.	Cell type	No.	Cell type	No.
anti-A	Normally blue coloured	A _X ,A ₃	A ₁	2	B	2	See insert of reference preparation(s)	
	Should equal or exceed potency of reference preparation (s)	A cord cells	A ₂ B	2	O	2		
	Should detect variants and subgroups as detailed in the manufacturer's instructions for use		A _X [*]					
anti-B	Normally yellow coloured	B _X , B ₃ , B _V	B	2	A ₁	2	See insert of reference preparation(s)	
	Should equal or exceed potency of reference preparation (s)	B cord cells	A ₁ B	2	O	2		
	Should detect variants and							

	subgroups as detailed in the manufacturer's instructions for use							
anti-A, B	Normally clear coloured	A ₁ , A ₂ , B, A ₁ B, A ₂ B	A ₁	1	O	4	See insert of reference preparation(s)	
	Should equal or exceed potency of reference preparation (s)	A _X A ₃	A ₂	2				
	Should detect variants and subgroups as detailed in the manufacturer's instructions for use	B _X B ₃	B	2				
		A and B cord cells	A _X	2				
anti-A1	Normally clear coloured		A ₁	2	A ₂	2	A ₁	2
	Should detect variants and subgroups as detailed in the manufacturer's instructions for use		A ₁ B	2	A ₂ B	2		
					B	2		
					O	2		
anti-D	Normally clear coloured	Weak D (500 sites /cell)	R ₁ r	2	r'r	1	See insert of reference preparation(s)	
	Should equal or exceed potency of reference preparation (s)		R ₂ r	2	r''r	1		
	Should detect variants and subgroups as detailed in the manufacturer's instructions for use	D ^{VI} type 1, D ^{VI} type 3, D ^{IV} , D ^V , D ^{VII} , DFR, DBT, R ₀ ^{Har}	Weak D*	2	rr	1		
anti-C	Normally clear coloured	C ^w , C ^x , r ^s	R ₁ r	1	R ₂ R ₂	1	R ₁ r	2
	Potency titre greater than 4 vs by	R ₂ R _Z	R ₁ R ₂	1 1	r''r	1		

	techniques detailed in manufacturer's instructions for use		R_1R_Z					
	Should detect variants and subgroups as detailed in the manufacturer's instructions for use		$r'r$	1	rr	1		
anti-E	Normally clear coloured	R_1R_Z	R_2r	1	R_1R_1	1	R_2r	2
	Potency titre greater than 4 vs by techniques detailed in manufacturer's instructions for use	E^W	R_1R_2	2	$r'r$	1		
	Should detect variants and subgroups as detailed in the manufacturer's instructions for use		$r''r$	1	rr	1		
anti-c	Normally clear coloured	$R_1R_Z, R_1^wR_1$	R_1r	2	R_1R_1	3	R_1r	2
	Potency titre greater than 4 vs by techniques detailed in the manufacturer's instructions for use		R_1R_2	1				
	Should detect variants and subgroups as detailed in the manufacturer's instructions for use		$r'r$	1				
anti-e	Normally clear coloured	R_2R_Z	R_2r	2	R_2R_2	3	R_2r	2
	Potency titre greater than 4 vs by techniques detailed in the manufacturer's instructions for use		R_1R_2	1				
			$r''r$	1				

	Should detect variants and subgroups as detailed in the manufacturer's instructions for use							
anti-C ^w	Normally clear coloured	R ₁ ^w R ₁ , r ^w r, R ₁ ^w r	R ₁ ^w r or	2	R ₁ r	1	R ₁ ^w r	2
	Potency titre greater than 4 vs by techniques detailed in the manufacturer's instructions for use		R ₁ ^w R ₂	2	R ₁ R ₁	1		
			r ^w r	1	r'r	1		
anti-K	Normally clear coloured	K+k+	K+k+	4	K-k+	4	K+k+	2
	Potency titre greater than 4 vs by techniques detailed in the manufacturer's instructions for use	Kp (a+b+)						
	Should detect variants and subgroups as detailed in the manufacturer's instructions for use	K+k+						
		Kp (a-b+)						
anti-k	Normally clear coloured	K+k+	K+k+	4	K+k-	4	K+k+	2
	Potency titre greater than 4 vs by techniques detailed in the manufacturer's instructions for use	Kp(a+)						
			Kp(a+b+)	2				
	Should detect variants and subgroups as detailed in the manufacturer's instructions for use		K+k+ Kp(a-)	2				
anti-Fy ^a	Normally clear coloured		Fy(a+b+)	4	Fy(a-)	4	Fy(a+b+)	2

	<p>Potency titre greater than 4 vs by techniques detailed in the manufacturer's instructions for use</p> <p>Should detect variants and subgroups as detailed in the manufacturer's instructions for use</p>							
anti-Fy ^b	<p>Normally clear coloured</p> <p>Potency titre greater than 4 vs by techniques detailed in the manufacturer's instructions for use</p> <p>Should detect variants and subgroups as detailed in the manufacturer's instructions for use</p>	Fy ^x	Fy(a+b+)	4	Fy(b-)	4	Fy(a+b+)	2
anti-Jk ^a	<p>Normally clear coloured</p> <p>Potency titre greater than 4 vs by techniques detailed in the manufacturer's instructions for use</p> <p>Should detect variants and subgroups as detailed in the manufacturer's instructions for use</p>		Jk(a+b+)	4	Jk(a-)	4	Jk(a+b+)	2
anti-Jk ^b	<p>Normally clear coloured</p> <p>Potency titre greater than 4 vs by</p>		Jk(a+b+)	4	Jk(b-)	4	Jk(a+b+)	2

	<p>techniques detailed in the manufacturer's instructions for use</p> <p>Should detect variants and subgroups as detailed in the manufacturer's instructions for use</p>							
anti-S	<p>Normally clear coloured</p> <p>Potency titre greater than 4 vs by techniques detailed in the manufacturer's instructions for use</p> <p>Should detect variants and subgroups as detailed in the manufacturer's instructions for use</p>	S+s-, S+s+, S-s+	S+s+	4	S-s+	4	S+s+	2
anti-s	<p>Normally clear coloured</p> <p>Potency titre greater than 4 vs by techniques detailed in the manufacturer's instructions for use</p> <p>Should detect variants and subgroups as detailed in the manufacturer's instructions for use</p>		S+s+	4	S+s-	4	S+s+	2
anti-M	<p>Normally clear coloured</p> <p>Potency titre greater than 2 vs by techniques detailed in the manufacturer's instructions for use</p>	M-N+ He+	M+N+	4	M-N+	4	M+N+	2

	Should detect variants and subgroups as detailed in the manufacturer's instructions for use							
anti-N	<p>Normally clear coloured</p> <p>Potency titre greater than 4 vs by techniques detailed in the manufacturer's instructions for use</p> <p>Should detect variants and subgroups as detailed in the manufacturer's instructions for use</p>		M+N+	4	M+N–	4	M+N+	2
anti-P1	<p>Normally clear coloured</p> <p>Potency titre greater than 4 vs by techniques detailed in the manufacturer's instructions for use</p> <p>Should detect variants and subgroups as detailed in the manufacturer's instructions for use</p>		P1+ strong	4	P1–	4	P1+	2
			P1+ weak	4				
anti-Le ^a	<p>Normally clear coloured</p> <p>Potency titre greater than 4 vs by techniques detailed in the manufacturer's instructions for use</p> <p>Should detect variants and subgroups as detailed in the</p>		Le(a+b–)	4	Le(a–)	4	Le(a+)	2

	manufacturer's instructions for use							
anti-Le ^b	<p>Normally clear coloured</p> <p>Potency titre greater than 4 vs by techniques detailed in the manufacturer's instructions for use</p> <p>Should detect variants and subgroups as detailed in the manufacturer's instructions for use</p>	A ₁ B Le(a–b+)	A ₁ B Le(a–b+)	4	Le(b–)	4	Le(b+)	2
'Other'	<p>Normally clear coloured</p> <p>Potency titre greater than 2 vs by techniques detailed in the manufacturer's instructions for use</p> <p>Should detect variants and subgroups as detailed in the manufacturer's instructions for use</p>		Heterozygous positive	4	Antigen negative	4	Heterozygous positive	2
* For reagents where reactivity against the antigen is claimed								

For batch acceptance testing the user must ensure that the typing reagent reacts with the weakest available antigen expressing cells (refer to batch release positive reactors in table 11.3) and does not produce false positives with cells negative for the antigen.

11.2.2: Anti-human globulin reagents

11.2.2.1: Introduction

Monoclonal antibodies have been developed which necessitate revision of the optimal composition of anti-human globulin reagents. For example, because of the limitations imposed by the presence of C3d on normal red cells, particularly in stored blood, conventional polyclonal anti-complement reagents rely on anti-C3c to detect *in vitro* bound complement and limited amounts of anti-C3d to detect *in vivo* bound

complement. However, some monoclonal IgM anti-C3d reagents can be used at concentrations adequate to detect both *in vitro* and *in vivo* bound complement without causing unwanted positive reactions with normal red cells and fresh, inert, group-compatible serum in routine tests.

11.2.2.2: General requirements

- anti-IgG is the essential component since the majority of red cell alloantibodies are non-complement binding IgG.
- anti-complement should be present in reagents recommended for use with serum test samples.
- anti-light chain activity is desirable in reagents recommended for use with plasma test samples in order to detect IgM antibodies at levels unable to be detected in direct agglutination tests, especially with washed red cells.
- anti-C4d must be avoided. It is accepted that very low titres of anti-C4c may occur in reagents of animal origin.
- Reagents should be tested for the presence of heterospecific antibodies which can cause haemolysis or agglutination of unsensitised red cells in the indirect antiglobulin test and for the presence of unwanted positive reactions.

11.2.2.3: Performance evaluation

Performance evaluation should be undertaken in accordance with:

- BS EN 13612:2002 Performance Evaluation of *In Vitro* Diagnostic Medical Devices.
- Reagents listed in Annex II, List A, of the EU *In Vitro* Diagnostic Medical Devices Directive must also comply with the Common Technical Specifications for *In Vitro* Diagnostic Medical Devices (2009/108 /EC).

Stability testing should be performed in accordance with:

- BS EN ISO 23640 *In vitro* diagnostic medical devices. Evaluation of stability of *in vitro* diagnostic Reagents.

11.2.2.4: Batch release testing requirements

Specificity testing

Tests for IgM or IgG red cell heterospecific antibodies

- Heterospecific antibodies can cause haemolysis or agglutination of unsensitised red cells in the indirect antiglobulin test. Details of tests for heterospecific antibodies are outlined in section 11.4.

Requirements

- The anti-human globulin reagent should not agglutinate or haemolyse washed unsensitised red cells from two individuals of group A1 RhD positive, two individuals of group B RhD positive and two individuals of group O RhD positive, whether or not treated with proteolytic enzyme (e.g. papain, bromelain or ficin).

Tests for unwanted positive reactions

- These test for excess anti-C3d and anti-C3c, which can cause unwanted positive reactions in the indirect antiglobulin test, and for the presence of any undesirable antibodies in the reagent. Details of tests are outlined in section 11.4.

Requirements

- All reactions should be negative on macroscopic examination.

anti-IgG potency: polyspecific anti-human globulin and anti-IgG reagents for use in tube or microplate techniques

- The anti-human globulin reference reagent should be tested in parallel with the test reagent, each being titrated against red cells sensitised with potent IgG anti-D antibody.

Requirements

- The potency titre of the test anti-human globulin or anti-IgG reagent should be at least equal to that of the reference reagent.

Potency tests

anti-IgG potency by chequerboard titration studies with red cells sensitised with weak IgG antibodies (anti-D, anti-K and anti-Fy^a)

- Test anti-human globulin or anti-IgG reagents against a selection of weak antibodies to determine the optimum potency. Antibody preparations should not be diluted and the use of single-donor antibody preparations is preferred. Antibodies should include:
 - an IgG anti-D to give an anti-human globulin potency titre of 8–32 using a pool of group O R₁r red cells from four individuals
 - an IgG to give an anti-human globulin potency titre of 8–32 using Kk red cells
 - an IgG anti-Fy^a, to give an anti-human globulin potency titre of 8–32 using Fy(a+b+) red cells.

Details of tests are outlined in section 11.4.

Requirements

- The anti-human globulin reagent or anti-IgG reagent is satisfactory if the reaction grade at all dilutions attains or exceeds that of the reference reagent without significant prozone, against red cells sensitised with all dilutions of the anti-D, anti-K and anti-Fy^a. In this context, a significant prozone is more than one grade difference between the reaction of the anti-human globulin reagent undiluted and 1 in 2.

anti-complement potency; polyspecific anti-human globulin reagents for use in tube tests

- Test anti-human globulin or anti-complement reagents against a selection of complement-coated red cells to determine the optimum potency. C3 and C4 complement-coated red cells should be prepared as described in section 11.4. In addition, anti-complement activity may be evaluated by tests with complement-fixing antibodies, such as anti-Jk^a.

Requirements

- The anti-human globulin reagent should have an anti-C4c titre of 1 in 2 or less.
- The anti-human globulin reagent should not affect a macroscopic reaction with EC4d red cells.

- The reagent should attain the potency titre of the reference reagent.
- Conventional (polyclonal) anti-human globulin or anti-human globulin containing monoclonal IgG anti-C3d that attain adequate reactivity with an optimal incubation period different from that recommended for the detection of IgG antibody, should state in the instructions for use the appropriate incubation period required for the optimum detection of red cell bound C3c/d complement components.

Tests for unwanted positive reactions

- These test for excess anti-C3d and anti-C3c, which can cause unwanted positive reactions in the indirect antiglobulin test, and for the presence of any undesirable antibodies in the reagent. Details of tests are outlined in section 11.4.
- All test results should be negative as defined by the manufacturer in the 'instructions for use'.

Instructions for use

The instructions for use for anti-human globulin reagents used in tube and microplate tests should also include a statement that:

- Inadequate washing of red cells in the anti-human globulin test may result in neutralisation of the anti-human globulin reagent.
- Following completion of the wash phase in the anti-human globulin test, excess residual saline may dilute the anti-human globulin reagent, when added, beyond that in the manufacturer's assessment.
- No single test is capable of detecting all clinically significant antibodies.
- For each batch of antibody screening being undertaken by an anti-human globulin test, a positive and negative control should be included. The positive control should be a weak anti-D (not more than 0.1 IU/mL); the negative control an inert serum, tested against the antibody screening cells being used.

11.2.3: Reagent red cells

11.2.3.1: Introduction

Reagent red cells prepared from human blood are essential in ensuring safe transfusion practice. They are used in the determination of ABO blood groups, in the control of blood grouping reagents and of the anti-human globulin technique, and in the detection and identification of atypical red cell alloantibodies.

11.2.3.2: General guidelines for reagent red cell manufacture

- When testing reagent red cells, in order to confirm the presence or absence of antigens listed in the antigen profile, a sample from each individual should be tested wherever possible, with a minimum of two antisera for each specificity prepared from different donors/cell lines.
- Where such testing produces conflicting results, repeat and further testing with at least one additional example of the relevant antibody(ies) should be undertaken to confirm the antigenic status of that cell.
- Where such testing has been performed with only one example of any blood grouping reagent, this information should be stated in the antigen profile included within the package insert.

- Reagent red cells should be shown not to produce unwanted positive reactions by the methods recommended for use by the manufacturer.
- Except for IgG-sensitised and C3-sensitised red cells, reagent red cells should be negative in the direct anti-human globulin technique with anti-IgG and polyspecific anti-human globulin reagents.
- With the exception of umbilical cord blood, alloabsorption and quantification cells, red cells used to test a patient's samples for atypical antibodies should not be pooled.
- Reagent red cells should be processed by a method and suspended in a medium that consistently ensures stability of the antigens specified in the antigen profile included within the package insert.
- With the exception of controls for automated systems representing whole blood, all red cell reagents should be free of ABH-specific blood group substances and blood group antibodies, including anti-A and anti-B, demonstrable by the manufacturer's recommended methods of use.
- The method of manufacture should ensure that white cells are removed from donations of red cells before the white cells lyse and release enzymes, which may adversely affect the properties of the red cells.

11.2.3.3: Immediate container label and/or instructions for use sheet

The immediate container and instructions for use sheet for reagent red cells should also meet the following criteria:

- Include a statement regarding the use of 'pooled cells', if cells are prepared from pooled material.
- Where reagent red cells are intended for use in ABO grouping or control of ABO or D blood grouping reagents, only the ABO and D group need be stated.
- When the reagent red cells are a multi-container product such as a red cell panel, the label on the immediate containers and packaging should be assigned the same identifying batch reference and carry a number or symbol to distinguish one container from another. This number or symbol should also appear in the antigenic profile.
- The date of expiry of reagent red cells should be stated on the antigenic profile.
- Where reagent red cells are provided suspended in preservative medium, the components of the medium should be stated in the instructions for use.
- The concentration and limits of the red cell suspension should be stated in the instructions for use.
- For enzyme-treated reagent red cells, information should be given in the instructions for use concerning those antigens which are rendered inactive or less active by the enzyme treatment used.

11.2.3.4: Reagent red cells for use in ABO and RhD grouping

- Reagent red cells should be groups A₁ and B. In addition, A₂ or O red cells may be included.
- At least one of the set should be RhD positive and one RhD negative.

11.2.3.5: Reagent red cells for use in antibody screening

The detection of atypical antibodies in the serum of a patient is of greater clinical significance than if such antibodies are detected in blood donors. Reagent red cells of a lesser specification may be used when performing antibody screening tests on blood donor samples.

In general the following should apply:

- Reagent red cells for use in antibody screening should be confirmed as group O by an ABO blood grouping procedure that is capable of demonstrating the A_x phenotype.
- Where practicable, reagent red cells known to express antigens having a frequency of less than 1% in the general population of the UK should not be included in reagent red cells for antibody screening.
- Where practicable, red cells from individuals known consistently to effect troublesome reactions with HLA antibodies should not be used as reagent red cells for antibody screening of patients.

11.2.3.6: Reagent red cells for use in antibody screening of patient samples

- As a minimum the following antigens should be expressed on the reagent red cells for antibody screening:

C, c, D, E, e, K, k, Fy^a, Fy^b, Jk^a, Jk^b, S, s, M, N, P₁, Le^a and Le^b

- As a minimum, reagent red cells from two individuals should be provided. These red cells should not be pooled. One reagent red cell should be R₂R₂; the other R₁R₁ (or R₁^wR₁).
- Apparent homozygous expression of the following antigens is desirable:

Fy^a, Fy^b, Jk^a, Jk^b, S and s

11.2.3.7: Reagent red cells for use in antibody screening of patient samples who have received prophylactic anti-D

- For pregnant patients who have received prophylactic anti-D, as a minimum the following antigens should be expressed:

c, e, K, k, Fy^a, Fy^b, Jk^a, Jk^b, S, s, M, N, P₁, Le^a and Le^b

- The cells must be RhD negative
- As a minimum, reagent red cells from two individuals should be provided. These red cells should not be pooled.
- Apparent homozygous expression of the following antigens is desirable:

Fy^a, Fy^b, Jk^a, Jk^b, S and s

11.2.3.8: Reagent red cells for use in antibody screening of donor samples

- Reagent red cells may be:
 - provided unpooled from a minimum of two individuals OR
 - as a pool of red cells in equal proportions from no more than two donors OR
 - red cells from a single donor.

- Pooled reagent red cells for antibody screening should be used only for testing samples from blood donors, not samples from patients.
- As a minimum the following antigens should be expressed:
D, C, c, E, e and K.
- To enhance the antigens of these screening cells they may be treated by proteolytic enzymes.

11.2.3.9: Reagent red cells for use in antibody identification

- Reagent red cells for use in the identification of atypical antibodies should be confirmed as group O by an ABO blood grouping procedure which is capable of demonstrating the A_x phenotype.
- Where practicable, red cells from individuals known consistently to effect troublesome reactions with HLA antibodies should not be used in reagent red cells for antibody identification.
- The antigen profile of reagent red cells for antibody identification should permit the identification of frequently encountered antibodies (e.g. anti-D, anti-E, anti-K and anti-Fya), and of commonly encountered alloantibody mixtures (e.g. anti-D+K).
- A red cell antibody identification panel comprises cells from eight or more individuals which should between them express the following antigens:

C, C^w, c, D, E, e, K, k, Kp^a, Fy^a, Fy^b, Jk^a, Jk^b, S, s, Le^a, Le^b, M, N, P₁ and Lu^a.

- Red cells from one individual should be R₁R₁ and from another R₁^wR₁ and between them should express the antigens:
K, k, Fya, Fyb, Jka, Jkb, S and s.
- Red cells from one individual should be R2R2, another r''r and those from another r'r.
- Red cells from a minimum of three individuals should lack the Rh antigens C, E and D. One of these three individuals should be K positive. Between them, red cells from these individuals should exhibit apparent homozygous expression of the antigens:

c, k, Fy^a, Fy^b, Jk^a, Jk^b, S and s.

11.2.3.10: Reagent red cells (IgG-coated) for use in the control of the anti-human globulin technique

- To ensure that the anti-IgG activity in negative antiglobulin tests has not been fully or partially neutralised, control red cells 'sensitised' with IgG antibody are added to negative tests.
- Group O RhD positive red cells are sensitised with sufficient anti-D to render an indirect antiglobulin test negative when a volume of these sensitised red cells and a volume of serum diluted 1 in 1000 are added, but remains positive if a volume of saline instead of diluted serum is added.

11.2.3.11: Reagent red cells for use in antibody in antibody strength determination (other than anti-D and anti-c)

Reagent red cells for use in the identification of atypical antibodies should be confirmed as group O by an ABO blood grouping procedure which is capable of demonstrating the A_x phenotype.

- As a minimum, reagent red cells from two individuals should be provided
- These red cells should not be pooled
- One D+ C+ E+ c- e+ (R₁R₂)
- One D+ C+ E+ c+ e+ (R₁R₂)
- Between them will show heterozygous expression of the following antigens: M, N, S, s, K, k, Fy^a, Fy^b, Jk^a and Jk^b
- They will be negative for Wr^a

11.2.3.12: Reagent red cells for use in patients with pan-reactive autoantibodies to determine the presence of underlying alloantibodies

Reagent red cells

- may be provided un-pooled OR as a pool of red cells from more than one donor
- should be selected to ensure the ID of underlying clinically significant antibodies to the following specificities;
 - C, C^W, c, D, E, K, Fy^a, Fy^b, Jk^a, Jk^b, S, s, M, N
- an R₂R₂ cell may be included to exclude the presence of an underlying allo anti-e in e negative patients
- Rh (D, C, E, c, e), K, Jk^a and Jk^b phenotypes must be provided for the users as a minimum.
- The cells may be enzyme treated which would result in them being negative for the following red cell antigens:
 - M, N, S, s, Fy^a and Fy^b

11.2.3.13: Reagent red cells for quantification of anti-D and anti-c antibody strength

Reagent red cells for use in quantification techniques should be confirmed as group O by an ABO blood grouping procedure which is capable of demonstrating the A_x phenotype.

Reagent red cells

- may be provided unpooled OR as a pool of red cells from more than one donor
- For anti-D quantification have the following phenotype: D+ C+ E- c- e+ (R₁R₁)
- For anti-c quantification have the following phenotype: D- C- E- c+ e+ (rr)
- The cells should be negative for C^W and K antigens
- The cells should be enzyme treated to enhance the antibody-antigen reaction»

11.2.3.14: Other reagent red cells

These reagent red cells should be manufactured in accordance with the general guidelines in section 11.2.3.2.

11.2.4: Miscellaneous reagents

11.2.4.1: Fetal calf serum and bovine serum albumin

When used in the formulation of reagents, fetal calf serum and bovine serum albumin should be obtained from a closed herd in the female line since 1980, in which no animal has been clinically suspected of having bovine spongiform encephalopathy (BSE), and which has not been fed rations containing ruminant-derived protein during that period.

Bovine albumin, usually supplied as a 20% or 30% solution, can be used as a constituent of a diluent for use in automated blood grouping antibody detection machines, for antibody quantification or as a potentiator in antisera, monoclonal reagents and anti-human globulin. When diluted and used in the system prescribed it should not cause:

- red cells to become T/Tk etc. transformed
- inhibition of antigen:antibody reactions
- false-positive reactions or rouleaux.

11.2.4.2: Proteolytic enzyme preparations

The activity of each batch of proteolytic enzyme should be assessed to ensure batch-to-batch consistency using a biochemical assay (e.g. azo-albumin technique).⁵

For manual antibody detection techniques, red blood cells treated with the enzyme should achieve activity comparable to that of the reference enzyme preparation 92/658 used with an anti-D of 2.5 to 3.5 IU/mL.

For automated antibody detection techniques for patient pre-transfusion samples red blood cells treated with the enzyme should readily detect a weak anti-D of no more than 0.1 IU/mL (e.g. NIBSC anti-D standard for assessing operator and test performance as described at www.nibsc.org).

For automated antibody detection techniques for donation testing the red blood cells treated with the enzyme should readily detect a weak anti-D of 0.5 IU/mL.

11.2.4.3: Water

The quality of water used in the production of a reagent should be adequate for that reagent. Ionic and non-ionic contaminants of water may interfere with components of reagents or may result in a conductivity or osmolality other than that intended. Water should have a conductivity of 1.0 μ S/cm or less or a resistivity of 1.0 Mohm/cm or greater.

11.2.4.4: Saline

Saline is an isotonic solution containing 8.5 to 9.0 g/L NaCl (0.145–0.154 M) and should contain sufficient buffer to maintain pH 7.0 \pm 0.2 at 22 \pm 1°C during its shelf life.

11.2.4.5: Low ionic strength solution

The term low ionic strength solution (LISS) should not be used to denote a low ionic strength formulation other than that described by Moore and Mollison.⁶ LISS should not be used in place of preparations designed for a particular technology. LISS has the following properties:

- pH 6.5–7.0 at 22 ±1°C
- conductivity 3.4–4.0 mS/cm
- osmolality 285–305 mOsmol/kg.

The reactions obtained by an indirect antiglobulin test (IAT) with a weak anti-D and D positive cells suspended in LISS should be equal to, or better than, those obtained with the same cells suspended in saline and incubated at 37°C for 15 minutes.

11.2.4.6: Weak antibodies for use as controls in antibody investigation techniques

Weak antibodies, such as anti-D, -K, -Fy^a can be used to control antibody detection techniques using indirect antiglobulin methods.

To act as a wash control the weak anti-D positive control could be diluted in serum or plasma. If the diluent is saline/bovine serum albumin, the control test could be positive, even though the cell washing was sub-optimal and this should be noted in the package insert.

These weak antibodies should:

- when used undiluted give a grade 2–4 reaction with red cells with homozygous antigen expression and have a mean IAT titre of 4 with the same cells.

For weak anti-D the antibody activity should be expressed in IU/mL.

11.2.4.7: Antibodies and cells representative of patient samples i.e. whole blood controls to control automated systems

As a minimum two vials containing red cells and plasma combined:

- One sample to be RhD positive, the other RhD negative with anti-D in the RhD negative plasma
- Red cells of the same ABO/D group may be pooled
- The red cells must give an unequivocal positive reaction with the appropriate ABO/Rh D grouping reagents
- One sample will contain anti-K or other non-Rh antibody (the red cells for this sample must be antigen negative for the corresponding antibody)
- Anti-D and anti-K (or other non-Rh antibody) must give an unequivocal positive reaction with antigen positive red cells by IAT at 37°C
- The plasma component of the whole blood controls should be free of other blood group antibodies unless stated in the instructions for use

11.2.4.8: AB Serum

The reagent should be:

- prepared from a pool of human group AB plasma or serum
- IAT antibody screen negative

- Negative for rouleaux inducing properties by direct agglutination at room temperature and by IAT and enzyme techniques at 37°C

11.2.4.9: Dithiothreitol (DTT)

DTT can be used to alter the red cell membrane and/ or reduce the disulphide bonds of IgM molecules and can be supplied at different concentrations to treat red cells and plasma samples or reagents.

11.2.4.10: Reagents for use in assessing the amount of D positive red cells in a suspected fetomaternal haemorrhage (FMH) by flow cytometry

Fluorescently labelled monoclonal antibodies used as a group of reagents to accurately determine the size of an RhD positive fetal bleed in an RhD negative person.

11.3: Reference preparations

11.3.1: Introduction

One of the major regulatory requirements is a requirement for traceability to reference materials of higher order. In the case of blood grouping reagents there are several national and international reference preparations already available to manufacturers to ensure adequate potency of anti-A, anti-B and anti-D grouping reagents and the potency and/or performance of a number of other serology reagents or procedures.

As batch identifiers may change during the lifetime of these guidelines please refer to www.nibsc.org for guidance.

11.3.2: International Standards for minimum potency of anti-A and anti-B blood grouping reagents

These anti-A and anti-B preparations are the lyophilised residues of culture supernatants from murine monoclonal hybridomas BRIC 131 and ES4 respectively. The preparations, when reconstituted and diluted according to the supplied instructions, define the minimum acceptable potency of manufactured anti-A, anti-B, anti-A,B and anti-A+B blood grouping reagents, i.e. the titre of the grouping reagent should be at least equal to that of the appropriate minimum potency reference preparation.

11.3.3: International Standard for minimum potency of anti-D blood grouping reagents for use in direct tests

This preparation is the lyophilised residue of culture supernatant from a human-murine monoclonal heterohybridoma secreting an IgM anti-D (RUM-1). When reconstituted and diluted according to the supplied instructions, this material defines the minimum acceptable potency of anti-D grouping reagents in direct tube tests, i.e. the titre of the grouping reagent should be at least equal to that of the minimum potency reference preparation in tube tests using unmodified red cells and without additional agents.

11.3.4: International Council for Standardization in Hematology/International Society of Blood Transfusion (ICSH/ISBT) reference preparations for papain and anti-D

The intended use of these preparations is to ensure adequate sensitivity combined with freedom from false-positive reactions associated with some manufacturers' enzyme preparations and techniques. The

recommended procedure is to test the papain reference material in conjunction with a suitable anti-D preparation of 2.5 to 3.5 IU/mL using a titration series for sensitivity, and a series of inert sera for false-positive reactions, according to the specified two-stage reference method in the product insert and to compare the titration scores with those obtained from testing the manufacturer's enzyme preparation in its recommended technique with the anti-D reference preparation and the inert sera.

11.3.5: UKBTS/NIBSC anti-D reference preparation for assuring operator and test performance

The current preparation (98/540) consists of lyophilised human plasma with a reconstituted anti-D potency of 1.8 IU/mL. At 1 in 20 dilution, it is intended to be used to assure the efficacy of red cell washing prior to the addition of an antiglobulin reagent. At 1 in 40 dilution, it is intended to be used in intra-laboratory monitoring to assess test operator variability in the detection of weak, macroscopic agglutination in the spin-tube antiglobulin test or equivalent reaction grades using automated methods.

11.4: Recommended serological techniques for reagent testing

11.4.1: Potency titrations

11.4.1.1: Introduction

The use of a semi-automatic pipette is recommended; one volume being in the order of 40 µL.

A separate pipette tip should be used for each reagent.

If the reagent is formulated with a medium to enhance its reactivity then the diluent for the determination of the potency titre should be a formulation identical to the reagent but with antibody protein replaced by non-antibody protein, e.g. fetal calf serum or bovine serum albumin. Otherwise, dilutions may be prepared in saline containing a final concentration of 20 g/L bovine serum albumin that has not been deliberately polymerised or otherwise potentiated.

Beginning with the undiluted blood grouping reagent, doubling dilutions (1 in 2, 1 in 4, 1 in 8 etc.) should be prepared. When preparing doubling dilutions, after the addition of the reagent or diluted reagent to an equal volume of the diluent, the tip of the pipette is emptied and blotted before the dilution is mixed and a volume transferred to prepare the subsequent dilution.

The potency titre is the reciprocal of the highest dilution of the reagent that effects a grade 2 reaction using tube and microplate or a grade 1 endpoint in column agglutination technologies.

The dilution caused by the addition of the cell suspension should not be considered in determining the potency titre.

11.4.1.2: Potency test methods for manual and microplate blood grouping reagents

Manual method – direct test

- Add one volume of each dilution of the reagent to a separate tube.
- Add one volume of 2–3% test red cell suspension to each tube.

- Mix thoroughly and incubate for the appropriate temperature and duration.
- Centrifuge and determine the reaction grade.

Manual method – indirect anti-human globulin test

- Add two volumes of each dilution of the reagent to a separate tube.
- Add one volume of 2–3% test red cell suspension in saline, or two volumes of 1.5–2% test red cell suspension in LISS.
- Mix thoroughly and incubate at 37°C for 45 minutes if the red cells are suspended in saline, or for 15 minutes if suspended in LISS.
- Wash the red cells four times.
- Add two volumes of anti-human globulin reagent to the button of test red cells. Mix. Centrifuge and determine the reaction grade.

Microplate method

Equipment

- Rigid polystyrene microplates with 'U'-shaped wells.
- Centrifuge with microplate carriers having a radius of at least 10 cm.
- Microplate shaker.
- Concave microplate reading mirror or automated plate reader.
- Red cells for microplate use, bromelain-treated if required.

Method

- Using a microplate, add one volume (25–50 μ L) of each dilution of the reagent to one volume of 2–3% test red cells.
- Mix the contents of the wells using a microplate shaker. Incubate at 19–25°C for 15 minutes.
- Centrifuge the microplate at 100g for 40 seconds. Gently dislodge the red cells from the bottom of the wells using a microplate shaker.
- Determine the reaction grade using a concave mirror or automatic plate reader.

11.4.1.3: Avidity determination

- Mix over an oval area of approximately 20 mm \times 40 mm on a glass slide, one volume of the undiluted reagent and one volume of a 30–45% red cell suspension in allogeneic serum or ABO group-compatible plasma.
- Maintain the slide at the recommended temperature for a slide test. If a range of incubation temperatures is given, for those blood grouping reagents where the antibody-antigen reaction is

favoured by a colder temperature, the higher temperature should be used; for other blood grouping reagents, the lower temperature should be used.

- Determine the time from mixing at which macroscopic agglutination first appears and record the reaction grade at 1 minute.

11.4.1.4: Test used in performance evaluation and batch release testing of anti-human globulin

Tests for IgM and IgG red cell heterospecific antibodies

- These test for heterospecific antibodies which can cause haemolysis or agglutination of unsensitised red cells in the indirect antiglobulin test.

Method

- Divide 12 test tubes into two sets of six.
- Into each of the first set of tubes, add one volume of washed 2–3% untreated red cells in saline from two group A₁ RhD positive, two group B RhD positive and two group O RhD positive individuals.
- Into each of the second set of tubes add one volume of washed 2–3% enzyme-treated red cells (papain, bromelain or ficin) in saline from the same group A₁ RhD positive, group B RhD positive and group O RhD positive individuals.
- Add two volumes of the anti-human globulin reagent, as intended to be supplied for use, to each test tube. Mix thoroughly. Incubate the reactants for five minutes at 19–25°C.
- Centrifuge the tubes.
- Determine the reaction grade.

Control of enzyme treatment

Weak IgG anti-D known to be reactive with enzyme-treated red cells should effect a positive reaction with each washed, enzyme-treated, red cell sample by the following method:

- To separate tubes, add one volume of the weak IgG anti-D to one volume of each of the washed, 2–3% suspension of enzyme-treated, RhD positive red cell samples. Mix thoroughly. Incubate for five minutes at 37°C. Centrifuge the tubes. Determine the reaction grade.
- The weak anti-D used for this purpose must be absorbed to remove anti-A or anti-B.
- Each of the enzyme-treated RhD positive red cell samples should be agglutinated by the weak IgG anti-D.

Tests for unwanted positive reactions

These test for excess anti-C3d and anti-C3c, which can cause unwanted positive reactions in the indirect antiglobulin test, and for the presence of any undesirable antibodies in the reagent.

Method for preparation of the red cell suspensions from segmented bleed line samples

- Select integral segment lines from two packs of group A1, two packs of group B and two packs of group O blood stored at 2–6°C for at least 10 days.

- Wash each of the red cell samples with saline sufficient to remove serologically reactive traces of plasma.
- Prepare suspensions of each red cell sample as 2–3% in saline and as 1.5–2% in LISS.

Incubation of red cells and fresh group-compatible serum

- Each of the six red cell samples described above is tested as a saline and a LISS suspension with a different, fresh, group-compatible serum.
- For each anti-human globulin reagent to be assessed, prepare two sets of six tubes.
- To the first tube of the first set of six tubes and the first tube of the second set of six tubes, add 1 mL of a fresh, single-donor group-compatible serum. Add 1 mL of a second fresh, single-donor group-compatible serum to the second tube of each set, and so on for the six different, fresh, group-compatible sera.
- To the first tube of the first set of six tubes, add 0.5 mL of a red cell sample as a 2–3% suspension in saline. Add 1 mL of the same red cell sample as a 1.5–2% suspension in LISS to the first tube of the second set of six tubes. Add 0.5 mL of the second red cell sample as a 2–3% suspension in saline to the second tube of the first set of tubes and 1 mL of the same red cell sample as a 1.5–2% suspension in LISS to the second tube of the second set of tubes, and so on for each of the six different, red cell samples.
- Incubate the first set of tubes (saline suspended red cell samples) for 45 minutes at 37°C. Incubate the second set of tubes (LISS suspended red cell samples) for 15 minutes at 37°C.
- Wash the red cell samples with saline sufficient to remove serologically reactive traces of serum. Resuspend the red cells to 2–3% in saline.

Tests with anti-human globulin reagents

- For each anti-human globulin reagent, prepare two sets of six tubes. To each of the first set of six tubes, add in sequence one volume of the 2–3% suspension of washed red cells from the saline test above.
- To each of the second set of six tubes, add in sequence one volume of the washed 2–3% suspension of washed red cells from the LISS tests above.
- Add two volumes of undiluted anti-human globulin, as supplied for use, to each of the 12 tubes. Mix thoroughly.
- Centrifuge the tubes.
- Determine the reaction grade.

anti-IgG potency: polyspecific anti-human globulin and anti-IgG reagents for use in tube or microplate techniques

The anti-IgG reference reagent (see section 11.3.5) should be tested in parallel with the test reagent, each being titrated against red cells sensitised with potent IgG anti-D.

Method

Test cells

- A 2–3% suspension in saline of washed pooled group O R₁r red cells is prepared from four individuals.

anti-D

- anti-D suitable for use in this application should have a potency titre of greater than 512.
- To 4 mL of the potent IgG anti-D add 2 mL of the 2–3% suspension of pooled group O R₁r red cells.
- Mix and incubate at 37°C for 45 minutes.
- Wash the red cell sample with saline sufficient to remove serologically reactive traces of serum. Prepare suspensions of each red cell sample as 2–3% in saline.

Technique

- Prepare 1 mL volumes of twofold serial dilutions of the test anti-human globulin reagent and anti-IgG reference preparation from 1 in 8 to 1 in 4096 (ten tubes).
- Prepare a set of ten tubes for each anti-human globulin reagent to be assessed.
- Place two volumes of each dilution into each of the series of ten tubes.
- Add one volume of the 2–3% suspension of pooled sensitised R₁r red cells to each tube, mix and centrifuge.
- Determine the potency titre.

Controls

The washed, strongly sensitised 2–3% suspension of R₁r red cells gives a negative result when centrifuged and gives negative results using the direct anti-human globulin technique with anti-complement (anti-C3c, anti-C3d, anti-C4c and anti-C4d) reagents and with anti-human globulin diluent in place of the anti-human globulin reagent. (The anti-complement specificities may be present as mixtures in one or more reagents.)

Test for anti-IgG potency by chequerboard titration studies with red cells sensitised with weak IgG antibodies (anti-D, anti-K and anti-Fy^a)

Selection of weak IgG antibody preparations

Antibody preparations should not be diluted to attain the following potency requirements. The use of single-donor antibody preparations is preferred.

The following are selected:

- an IgG anti-D to give an anti-human globulin potency titre of 8–32 using a pool of group O R₁r red cells from four individuals
- an IgG anti-K containing a final concentration of 0.014M EDTA neutralised to pH 7, to give an anti-human globulin potency titre of 8–32 using K+k+ red cells
- an IgG anti-Fy^a containing a final concentration of 0.014M EDTA neutralised to pH 7, to give an anti-human globulin potency titre of 8–32 using Fy(a+b+) red cells.

Test cells

Prepare 10 mL of a 2–3% suspension of washed R₁r red cells pooled in equal proportions from four individuals. Similarly, prepare 10 mL of a 2–3% suspension of washed Kk red cells and 10 mL of a 2–3% suspension of washed Fy(a+b+) red cells.

Sensitisation of test cells

anti-D

- Using a set of five containers each of 20 mL to 25 mL volume, prepare 4 mL volumes of serial twofold dilutions of the anti-D from undiluted to 1 in 16.
- Add 2 mL of the 2–3% suspension of pooled R₁r red cells in saline to each container. Mix and incubate at 37°C for 45 minutes.
- Wash the red cells four times with 20 mL volumes of saline at each wash and remove the last supernatant.
- Add 2 mL of saline to the packed washed red cells to prepare the 2–3% suspensions of sensitised red cells.

anti-K

As above, but using the anti-K with the K+k+ red cells.

anti-Fy^a

As above, but using the anti-Fy^a, with the Fy(a+b+) red cells.

Preparation of anti-IgG and/or anti-human globulin dilutions

For each anti-IgG and/or anti-human globulin under test and the anti-IgG reference preparation, prepare 2 mL volumes of twofold serial dilutions from undiluted, that is as supplied for use, to 1 in 16.

Test method for anti-IgG or antiglobulin potency by chequerboard titration

anti-D sensitised red cells

- Prepare five sets of five tubes for each anti-human globulin reagent under test and the anti-IgG reference reagent.
- Place two volumes of the anti-human globulin reagent, undiluted to 1 in 16, in the appropriate tubes for each of the five sets of five tubes.
- Using the 2–3% suspension of red cells sensitised with the undiluted anti-D for the first set of five tubes, the 2–3% suspension of red cells sensitised with the anti-D diluted 1 in 2 for the second set of five tubes, and so on, finishing with the 2–3% suspension of red cells sensitised using the anti-D diluted 1 in 16 for the fifth set of five tubes, add one volume of the washed red cells to each of the sets of anti-human globulin dilutions (see Table 11.4).
- Mix thoroughly. Centrifuge the tubes, appropriately.
- Determine the reaction grade.

Table 11.4 Chequerboard test format

Set	anti-D used to coat red cells	Dilution of anti-human globulin reagent				
		N	2	4	8	16
1	Undiluted					
2	1 in 2					
3	1 in 4					
4	1 in 8					
5	1 in 16					

anti-K sensitised red cells

As above, but using the anti-K sensitised K+k+ cells.

anti-Fy^a sensitised red cells

As above, but using the anti-Fy^a sensitised Fy(a+b+) cells.

Controls

The unwashed 2–3% red cell suspensions sensitised with the undiluted anti-D, anti-K and anti-Fy^a give negative results in a spin-tube test. The washed sensitised cells should not react with the diluent or the anti-complement components of the anti-human globulin reagents.

*Test for anti-complement potency; polyspecific anti-human globulin reagents for use in tube tests⁷***Preparation of the complement sensitised red cells**

Various very low ionic strength medium techniques are used to prepare the iC3b, C4b, C3d and C4d sensitised red cells that are necessary for the assessment of anti-complement activity.

The C3 and C4 activation states produced on red cells by the various methods are shown in Table 11.5.

As a minimum, red cell samples from two individuals are to be prepared and tested as described below.

anti-C4b potency*Method*

- Prepare a set of three tubes for each anti-human globulin reagent under test.
- Prepare doubling dilutions of the anti-human globulin reagent from undiluted to 1 in 4.

- Place two volumes of each anti-human globulin dilution in the appropriate tubes.
- Add one volume of 2–3% EC4b red cells to each tube. Mix thoroughly. Centrifuge the tubes.
- Determine the reaction grade.

Controls

The EC4b cells do not react with anti-C3c, anti-C3d, anti-IgG or saline or the inert anti-human globulin diluent using the direct anti-human globulin technique. They react with anti-C4c and anti-C4d reagents.

anti-C4d potency

Method

- Place two volumes of undiluted anti-human globulin in a tube.
- Add one volume of 2–3% EC4d red cells. Mix thoroughly. Incubate for 5 minutes at 19–25°C.
- Centrifuge the tubes. Determine the reaction grade.

Controls

The EC4d cells do not react with anti-C3c, anti-C3d, anti-C4c, anti-IgG or saline or the inert anti-human globulin diluent using the direct anti-human globulin technique. The undiluted anti-human globulin does not agglutinate unsensitised red cells that have been trypsin-treated, using the direct anti-human globulin technique.

anti-C3d potency

Method

- Prepare a set of seven tubes for each anti-human globulin under test and the anti-C3d reference reagent (see section 11.3.5) which is tested in parallel, at the dilution for the 'immediate test' stated in its accompanying instructions for use.
- Place two volumes of each anti-human globulin dilution in each of the tubes (undiluted, that is as intended to be supplied for use, to 1 in 64).
- Add one volume of the 2–3% EC3d/EC4d red cells to each tube. Mix thoroughly and centrifuge the tubes, appropriately.
- Determine the reaction grade.

Controls

The EC3d/EC4d cells do not react with anti-C3c, anti-C4c, anti-IgG, saline or anti-human globulin diluent using the direct anti-human globulin technique. They do react with anti-C3d.

Table 11.5 Complement C3 and C4 activation

Method of preparation	Initial state	State after trypsin treatment

Very low ionic strength medium* 37°C	iC3b/C4b	iC3d/C4d
Very low ionic strength medium* 37°C	C3dg	C3d
Very low ionic strength medium* 37°C with EDTA	C4b	C4d
* These media are not to be confused with low ionic strength solution (LISS).		

11.5: References

1. Directive 98/79/EC of the European Parliament and of the Council of 27 October 1998 on *in vitro* diagnostic medical devices. OJ, L 331, 07.12.1998, p1.
2. Medicines and Healthcare products Regulatory Agency. *Rules and Guidance for Pharmaceutical Manufacturers and Distributors*. London: Pharmaceutical Press.
3. The Human Tissue Act 2004. Available at www.legislation.gov.uk
4. The retention and storage of pathological records and specimens (5th edition). Guidance from The Royal College of Pathologists and the Institute of Biomedical Science.
5. Phillips PK, Prior D, Dawes BA (1984). Modified azo-albumin technique for the assay of proteolytic enzymes for use in blood group serology. *Journal of Clinical Pathology*, 37, 329–331.
6. Moore HC, Mollison PL (1976). Use of low ionic strength medium in manual tests for antibody detection. *Transfusion*, 16, 291.
7. Lachmann PS, Voak D, Oldridge RG, Downie RM, Bevan PC (1983). Use of monoclonal anti-C3 antibodies to characterise the fragments of C3 that are found on erythrocytes. *Vox Sanguinis*, 45, 367–372.